

357 batches produced
(100 %)

11 technical failure (3.1 %)

- 1 human error
- 10 equipment failure

6 chemical failure (1.7 %)

- 6x ^{68}Ga -colloids > 3 %
- 5x ^{68}Ga -ions > 2 %
- 2x pH < 6.5

340 batches released
(95.2 %)

875 patient dosages
administered and PET/CT
scans performed