



READY

SOP

(none)

Load SOP...

Sample Information

File ID: Test

Sample ID:

Edit Info...

SOM

Control Mode: Total Count
125000Aperture: 100 μ m

Kd: 126.39

Current: 1600 μ A

Gain: 2

Runs: 1

Wait Time: 0 s

Save: Yes

Print: No

Export: Yes

Edit SOM...

Stirrer: Off 9.4°

Waste
TankWaste
JarElectrolyte
Jar

Create Preferences Wizard - Printed Report

☒ Sample Info☐ Short☐ Filtration Efficiency (Requires 2
overlaid runs)

Sizes...

Step 1 of 9

Size

☒ Graphs

Select...

☒ Statistics☒ Averaged Statistics☐ Compare to Sample Specifications☒ Overlay Statistics

Select...

☐ Listing

Select...

☐ Interpolation

Select...

Size Trend

☐ Graph

Select...

☐ Listing

Select...

Size Statistics

☐ Mass☐ Number☐ %☐ /ml☒ Volume☒ %☐ /ml☐ Surface Area☐ %☐ /ml

Pulse Data

☐ Graph

Select...

☐ Listing

Select...

Blockage Monitors

☐ Graph

Select...

☐ Listing

Select...

Printer

☒ Use Default Printer

Select...

☐ Use Specified Printer:

Clear All

Report Printing Order

Sample Info
Size Graphs
Size Statistics
Average Statistics
Overlay Statistics

First

Move Up

Move Down

Last

< Back

Next >

Cancel



Preview



Start



Repeat



Pause



Resume



Continue



Cancel



Stop



Reset



Unblock



Empty



Flush



Fill