

Materials List for:

Method for Simultaneous fMRI/EEG Data Collection during a Focused Attention Suggestion for Differential Thermal Sensation

Pamela K. Douglas^{1,2}, Maureen Pisani², Rory Reid¹, Austin Head², Edward Lau², Ebrahim Mirakhor³, Jennifer Bramen², Billi Gordon², Ariana Anderson², Wesley T. Kerr², Chajoon Cheong⁴, Mark S. Cohen^{1,2}

¹Neuropsychiatric Institute, University of California, Los Angeles

²Laboratory of Neuroimaging Technology, University of California, Los Angeles

³Yale School of Medicine

⁴Korean Basic Science Institute

Correspondence to: Pamela K. Douglas at PDouglas@mednet.ucla.edu

URL: <https://www.jove.com/video/3298>

DOI: [doi:10.3791/3298](https://doi.org/10.3791/3298)

Materials

Name	Company	Catalog Number	Comments
Precision Monolithic Temperature Sensor	National Semiconductor Corporation	LM34/LM35	100 mV/°C; accurate to within ±0.4 °C at room temperature
Precision Instrumentation Amplifier	National Semiconductor Corporation	INA114	Part of Magnet Room I/O interface
HydroCel Saline	Electrical Geodescis Inc.	N-PRT-KCL-1000-000	
HydroCel Geodesic Sensor Net with a 256-channel High Density Electrode Array	Electrical Geodescis Inc.	256-channel HCGSN	
Geodesics Photogrammetry System	Electrical Geodescis Inc.	EGI GPS	
Pipettes	Electrical Geodescis Inc.	N-ACC-PIP-1000-000	