

Video Article

A Detailed Protocol for Perspiration Monitoring Using a Novel, Small, Wireless Device

Kazuhiro Ogai¹, Masakazu Fukuoka², Kei-ichiro Kitamura³, Kiyoshi Uchide⁴, Tetsu Nemoto³

¹Wellness Promotion Science Center, Institute of Medical, Pharmaceutical and Health Sciences, Kanazawa University

²Advanced Research Center for Human Sciences, Waseda University

³Department of Clinical Laboratory Science, Graduate School of Medical Science, Kanazawa University

⁴Asanogawa General Hospital

Correspondence to: Kazuhiro Ogai at kazuhiro@staff.kanazawa-u.ac.jp

URL: <https://www.jove.com/video/54837>

DOI: [doi:10.3791/54837](https://doi.org/10.3791/54837)

Keywords: Medicine, Issue 117, Perspiration monitoring, water vapor flux, sensor, silica gel, sympathetic activity, emotional sweating, data processing, transepidermal water loss

Date Published: 11/24/2016

Citation: Ogai, K., Fukuoka, M., Kitamura, K.i., Uchide, K., Nemoto, T. A Detailed Protocol for Perspiration Monitoring Using a Novel, Small, Wireless Device. *J. Vis. Exp.* (117), e54837, doi:10.3791/54837 (2016).

Abstract

Perspiration monitoring can be utilized for the detection of certain diseases, such as thermoregulation and mental disorders, particularly when the patients are unaware of such disorders or are having difficulty expressing their symptoms. Until now, several devices for perspiration monitoring have been developed; however, such devices tend to have a relatively large exterior, considerable power consumption, and/or less sensitivity.

Recently, we developed a small, wireless device for perspiration monitoring. The device consists of a temperature/relative humidity (T/RH) sensor, battery-driven small data logger, and silica gel as a desiccant in a small cylindrical exterior. The T/RH sensor is placed between the detection windows (through which the water vapor from the skin enters) and the silica gel. The underlying principle of the perspiration monitoring device is based on Fick's law of diffusion, which means that water vapor flux from the skin to the silica gel (*i.e.* transepidermal water loss and perspiration) can be captured by change in humidity at the T/RH sensor. In addition, a baseline subtraction method was adopted to distinguish perspiration and transepidermal water loss.

As shown in the previous report, the developed device can monitor the perspiration at any sites of the body in an easy, wireless manner. However, detailed methods of how to use the device have not been disclosed yet. In this article, therefore, we would like to show the point-by-point tutorials of how to use the device for perspiration monitoring, by showing the sympathetic activity test with the sympathetic skin response monitoring as an example.

Video Link

The video component of this article can be found at <https://www.jove.com/video/54837/>

Introduction

Human perspiration, generally known as "sweating," is not just a mechanism for thermoregulation¹, but it is also related to certain kinds of diseases. The etiology of abnormal perspiration is broad, including: heatstroke, hyper- or hypothyroidism², brain infarction³, diabetes mellitus⁴, dysautonomia⁵, menopause (known as "hot flash")⁶, cystic fibrosis⁷, Parkinson's disease⁸, and social anxiety disorder⁹. In light of the number of perspiration-related diseases, it has been considered beneficial to monitor perspiration rates for the early diagnosis or prediction of such diseases (*e.g.*, prevention of heatstroke) in a ubiquitous manner¹⁰.

To date, only a small number of devices for perspiration monitoring have been proposed. In early days, skin conductance and relative humidity were used for indirect indices of the amount of perspiration^{11,12}. Most recently, several kinds of flexible, wearable sensors for perspiration monitoring have been proposed¹³⁻¹⁹, although they are intended for the analysis of sweat electrolytes rather than the amount or temporal pattern of perspiration. The calculation of water vapor diffusion has been utilized for a more quantitative method of monitoring water exchange from the skin²⁰⁻²³. However, this requires (1) the assumption that the outer atmosphere is still and constant²⁰, (2) enough sensitivity to detect the natural flow of water vapor^{21,22}, or (3) a coolant (*e.g.*, Peltier device that consumes a substantial amount of electricity) to condense water vapor to liquid²³; thus, they might be difficult for daily and long-term monitoring. As an alternative, a ventilated chamber method was developed^{20,24,25}. In the ventilated chamber method, dry nitrogen or dehumidified air is infiltrated in a small chamber adjacent to the skin from a nitrogen gas tank or a pump, and gas in the water vapor evaporated from the skin is collected. The amount of water vapor from the skin can be calculated from the difference of the humidity in the outlet and inlet gases. Although this method can estimate the amount of perspiration very precisely, a nitrogen gas tank or a mechanical pump is generally large enough to impede daily monitoring.

To address these drawbacks, we have recently developed a novel device for perspiration monitoring, in which a closed chamber with a desiccant-driven enforced water vapor flow, enabled sensitive and long-term monitoring²⁶. This device consists of a cylindrical plastic exterior, temperature/relative humidity (T/RH) sensor with recording microprocessor, and silica gel (**Figure 1**). In principle, the outer atmosphere should not interfere with the water vapor flow, and a coolant or ventilating chamber is not required. Perspiration profiles can be obtained by solving equations using a spreadsheet software²⁶. A previous study has only shown the principle of the developed device and has omitted the detailed method for how to use the device because of space limitations.

The objective of this article, therefore, is to show a detailed method of how to use the developed device for perspiration monitoring, by showing the recording of stress-induced palmar perspiration during the sympathetic activity test as an example.

Protocol

NOTE: The device, including the method of analysis, is covered by Japanese Unexamined Patent Application Publication No. 2011-169881 and the Japanese Patent No. 5708911. This study, including the protocol of the experiment with human subjects, was approved by the Medical Ethics Committee of Kanazawa University (#553-1).

1. Prerequisites for the Perspiration Monitoring Device

NOTE: Perform these steps only once before the first use.

1. Install the universal serial bus (USB)-serial port conversion interface drivers in the computer²⁷. If the drivers are already installed, skip this step.
2. Connect the USB-serial port conversion interface to the computer via USB cable. Wait for the automatic driver installation if any.
3. Set the parameters of the conversion interface and check the ID number of the serial port as follows.
 1. Open the device manager, and select [View] → [Devices by Type].
 2. Locate the "Ports (COM & LPT)" heading, and expand the section to find the heading containing "CP210x USB to UART Bridge" in the list.
 3. Memorize the ID number of the port (e.g., "COM5") in the "CP210x USB to UART Bridge" heading.
 4. Double click the "CP210x USB to UART Bridge," and open the "Port Settings" tab.
 5. Set the parameters as follows: "Bits per second" = 9600, "Data bits" = 8, "Parity" = "None," "Stop bits" = 1, and "Flow control" = "None."
 6. Click the "OK" button to close the window, and close the device manager.

2. Setup of the Perspiration Monitoring Device

NOTE: Set the recording settings as follows before using the device. Repeat these steps for each device if multiple devices are to be used.

1. Make sure the battery is not inserted in the perspiration monitoring device.
2. Plug the white connector of the USB-serial port conversion interface to the receptacle at the top of the perspiration monitoring device.
3. Execute the perspiration recording software, followed by setting the device parameters as follows.
 1. Open the "Settings" tab.
 2. Set the "COM" number to the ID number of the serial port (see step 1.3.3).
 3. Click the "Confirm Connection" button. Check if the message "Connected..." appears.
 4. At the "Data File Folder" settings, choose the drive and select the folder where the perspiration data will be saved. To select the subfolder, double click the folder to open.
 5. Set the "Number of holes" to "4."
4. Open the "Measure & Rec" tab, and set the "Time Interval" to the desired sample time.
5. Insert a battery to the perspiration monitoring device.
6. Fill the device with dry silica gel (color should be blue or green if the color indicator is available), and close the lid. If the lid does not close completely, reduce the amount of silica gel.
7. Click the "Start Logging Now" button in the perspiration recording software.

NOTE: Because the perspiration recording starts just after the disconnection, the perspiration monitoring device may be left connected until the monitoring is to be started.

3. Setup for the Measurement of Sympathetic Skin Response (SSR)

NOTE: These steps are intended to monitor the sympathetic activities as the palmar SSR, and are not necessarily required for perspiration monitoring itself. SSR is a change of skin potential according to the sympathetic arousal stimulation such as upset and concentration^{28,29}.

1. Cleanse the skin where the SSR is to be recorded with an alcohol swab.
2. Put the anode, cathode, and grounding electrode on the palm, back of the hand, and wrist, respectively, by means of an electrode paste. Fix the electrodes with medical tape.
3. Set the conditions of the instrumentation amplifier by turning the corresponding knobs of the amplifier as follows: Sensitivity = 1 mV/V, High cut filter ("HI CUT") = 3 kHz, and Low cut filter ("LO CUT") = 0.5 Hz.
4. Execute the SSR recording software, and start the recording at a sampling rate of 200 Hz by clicking the "Start Measurement" button in the software.

4. Recording of the Perspiration

1. Put a medical double-sided tape on the bottom of the perspiration monitoring device. When putting a tape, make sure the measurement windows (*i.e.* four holes at the bottom of the device) are not obstructed.
2. Disconnect the perspiration monitoring device from the USB-serial port conversion interface. Immediately after the disconnection, observe the perspiration monitoring begin automatically. Make sure the LED lamp is blinking.
3. Remove the release liner of the medical double-sided tape, and put the perspiration monitoring device on the skin where perspiration is to be monitored.
4. Wait 10 min for stabilization of water vapor diffusion.
NOTE: Although this process cannot be monitored, the previous study has confirmed that waiting 10 min is sufficient for stable monitoring²⁶.
5. Start the examination (*e.g.*, sympathetic activity test).
6. After the examination, remove the perspiration monitoring device from the skin. To stop the SSR recording, click the "Stop Measurement" button in the SSR recording software, and remove all electrodes from the skin.

5. Perspiration Analysis

1. Connect the perspiration monitoring device to the USB-serial port conversion interface (as in step 2.2).
2. Execute the perspiration monitoring software.
3. Open the "Settings" tab, and click the "Confirm Connection" button.
4. Open the "Measure & Rec" tab, and click the "Download" button to save the raw data of the device in the specified folder (see step 2.3.4).
5. Perform perspiration analysis as follows.
 1. Open the "Analysis" tab, and click the "Read Data File." Observe the "Open File" dialog pop up. Select the saved file in the step 5.4, and then click the "Open" button.
NOTE: The software automatically performs perspiration analysis based on the published method²⁶, and the results will appear on the screen. At the same time, the perspiration data as the comma-separated values format will be saved in the same folder.

Representative Results

Using this novel device for perspiration monitoring (**Figure 1**) and the Fick's law-based calculation, the temporal perspiration profiles can be obtained in an easy, wireless manner. **Figure 2** shows representative data of wireless perspiration monitoring during the sympathetic activity test. In the experiment, the device for the perspiration monitoring, along with the electrodes for the sympathetic skin response (SSR) monitoring, was attached to the subject's palm. For the sympathetic activity test, the subject was requested to sit and perform the following tasks: (1) take a deep inspiration 5 times with 1 min intervals, and (2) do a mental calculation (*e.g.*, continuously subtract 7 from 100, or the "Flash Anzan" in which the participant sums up the number displayed one after the other on a computer screen) to evoke sympathetic activity. The perspiration and the SSR were simultaneously recorded during the stress conditions. As a result of deep inspiration and mental calculation, the sympathetic activity-induced palmar perspiration could be measured using the developed device. **Figure 3** shows representative data of multipoint measurement during daily activities. About 1 hr recording of perspiration at the palm (emotional sweating) and the anterior chest (thermal sweating) showed distinct patterns according to the activities.




Figure 1: The Novel Device for Perspiration Monitoring. (A) The exterior of the device containing dry silica gel. (B, C) A "doughnut-shaped" double-sided tape used in this study and its attachment to the device. (D) Attachment of the device to the skin. [Please click here to view a larger version of this figure.](#)




Figure 2: Representative Recording of Perspiration Under the Sympathetic Activity Test. (A) Attachment of the perspiration monitoring device with the sympathetic skin response (SSR) electrodes on the palm. (B) As a result of sympathetic activity test, palmar perspiration along with the SSR reaction was observed in response to the sympathetic activities. [Please click here to view a larger version of this figure.](#)




Figure 3: Representative Multipoint Recording of Perspiration During the Daily Activities. The distinct patterns of perspiration according to the different parts of the body (red line, at the palm; blue line, at the anterior chest) were observed during several activities (e.g., going downstairs, driving a car, having a lunch with talking, and shopping). [Please click here to view a larger version of this figure.](#)

Discussion

The aim of this article is to introduce the use of a novel, wireless perspiration monitoring device. Owing to the recent progress of engineering, more accurate, easy-to-handle methods for temporal perspiration monitoring have been proposed; the ventilated chamber method^{24,25} and the vapor pressure diffusion method²³ are representative examples. However, the ventilated chamber method requires the use of dry nitrogen or a pump with desiccant to produce a dry atmosphere, and thus, the exterior tends to be large. The vapor pressure diffusion method can be adopted in a small exterior, although the open chamber system can be largely influenced by the outer atmosphere, and the closed chamber system has a problem with either water vapor saturation (the chamber is filled with a saturated water vapor) or considerable power consumption of a water vapor condenser (e.g., Peltier device).

To address the situation, we have recently developed a novel wireless device for perspiration monitoring²⁶. In the device, the water vapor from the skin can be captured by a desiccant, enabling a constant but natural flow. In the course of the water vapor traveling, the T/RH sensor captures the flow of the water vapor as a change in the temperature and relative humidity. By calculating such changes under Fick's law, we could estimate the temporal changes of perspiration amount (mg/cm²/min) with <5% error, relative to the conventional ventilated chamber method²⁶. The developed device can be used, for example, to monitor mental stress (**Figure 2**), thermal sweating, and ultimately the perspiration-related dysregulation in an easier manner.

The developed device is easy to handle, and so there are few points to consider. However, the perspiration measurement would fail if the silica gel is not dry. Therefore, before the experiment, the examiner should make sure that the silica gel is completely dry (*i.e.* the color is blue). A previous study has estimated the maximum duration of measurement being longer than 4 hr without changing the silica gel²⁶. To re-nature silica gel, dry it in a drying oven until the color turns blue. A conventional microwave oven is also useful. The perspiration measurement would also fail if the attachment of the device is not sufficient. We recommend the usage of a "doughnut-shaped" double-sided tape.

Nonetheless, the developed device has a limitation. Because of the delicate calculation and the limit of desiccant absorptivity, the error of the amount of perspiration should be considered, especially at the higher level of perspiration. Although we have confirmed the error rate as being <5% relative to the conventional method²⁶, the absolute value of the calculated perspiration value should be handled with caution.

Owing to the small, simple, and wireless design of the device, the multipoint perspiration monitoring under unrestricted conditions can be possible. As shown in **Figure 3**, the different temporal profiles of perspiration can be detected on the different positions of the skin (*i.e.* palm and anterior chest) in daily life conditions (e.g., talking, eating foods, driving a car, shopping, *etc.*). The device, therefore, might be utilized for the simultaneous detection of abnormal sweating derived both from the mental and thermal regulatory system malfunction. For example, the dysregulation of the peripheral nervous system in diabetes might be detected if the unbalanced sweating between the left and right soles were to be detected. We are now planning an observational study of perspiration profiles in hospitalized patients as a clinical experiment.

Disclosures

The author Fukuoka Masakazu is the president of the Rousette Strategy Inc., where the developed device was assembled. No financial support was received from either Fukuoka Masakazu or the Rousette Strategy Inc.

Acknowledgements

We would like to acknowledge Mr. Ryohei Suganuma and Ms. Sakie Tachibana who helped perform the research. This study was supported in part by the JSPS KAKENHI Grant Numbers 15K20664, 24500848, and 21500405. This study was also funded in part by the MEXT/JST Tenure Track Promotion Program. A part of this study was based on the Japanese Unexamined Patent Application Publication No. 2011-169881 and the Japanese Patent No. 5708911.

References

- Hardy, J. D. Physiology of temperature regulation. *Physiol. Rev.* **41**, 521-606 (1961).
- Niepomniszcze, H., & Amad, R. H. Skin disorders and thyroid diseases. *J. Endocrinol. Invest.* **24** (8), 628-638 (2001).
- Korpelainen, J. T., Sotaniemi, K. A., & Myllyla, V. V. Hyperhidrosis as a reflection of autonomic failure in patients with acute hemispherical brain infarction. An evaporimetric study. *Stroke.* **23** (9), 1271-1275 (1992).
- Fealey, R. D., Low, P. A., & Thomas, J. E. Thermoregulatory sweating abnormalities in diabetes mellitus. *Mayo Clin. Proc.* **64** (6), 617-628 (1989).
- Leung, A. K., Chan, P. Y., & Choi, M. C. Hyperhidrosis. *Int. J. Dermatol.* **38** (8), 561-567 (1999).
- Kronenberg, F., Cote, L. J., Linkie, D. M., Dyrenfurth, I., & Downey, J. A. Menopausal hot flashes: thermoregulatory, cardiovascular, and circulating catecholamine and LH changes. *Maturitas.* **6** (1), 31-43 (1984).
- Gibson, L. E., & Cooke, R. E. A test for concentration of electrolytes in sweat in cystic fibrosis of the pancreas utilizing pilocarpine by iontophoresis. *Pediatrics.* **23** (3), 545-549 (1959).
- Swin, L. *et al.* Sweating dysfunction in Parkinson's disease. *Mov. Disord.* **18** (12), 1459-1463 (2003).
- Davidson, J. R., Foa, E. B., Connor, K. M., & Churchill, L. E. Hyperhidrosis in social anxiety disorder. *Prog. Neuropsychopharmacol. Biol. Psychiatry.* **26** (7-8), 1327-1331 (2002).
- Gun Park, D., Chul Shin, S., Won Kang, S., & Tae Kim, Y. Development of flexible self adhesive patch for professional heat stress monitoring service. *Conf. Proc. IEEE Eng. Med. Biol. Soc.* **4**, 3789-3792 (2005).
- Pickup, J. C. Preliminary evaluation of a skin conductance meter for detecting hypoglycemia in diabetic patients. *Diabetes Care.* **5** (3), 326-329 (1982).
- Stenstrom, S. J. A study on skin humidity in leprosy patients using a new type of humidity meter. *Int. J. Lepr. Other Mycobact. Dis.* **52** (1), 10-18 (1984).
- Monty, C. N., Wujcik, E. K., & Blasdel, N. J. *Flexible Electrode for Detecting Changes in Temperature, Humidity, and Sodium Ion Concentration in Sweat.* US patent US 2013/0197319 A1 (2013).
- Wujcik, E. K., Blasdel, N. J., Trowbridge, D., & Monty, C. N. Ion Sensor for the Quantification of Sodium in Sweat Samples. *IEEE Sens. J.* **13** (9), 3430-3436 (2013).
- Blasdel, N. J., Wujcik, E. K., Carletta, J. E., Lee, K. S., & Monty, C. N. Fabric Nanocomposite Resistance Temperature Detector. *IEEE Sens. J.* **15** (1), 300-306 (2015).
- Jia, W. *et al.* Electrochemical tattoo biosensors for real-time noninvasive lactate monitoring in human perspiration. *Anal. Chem.* **85** (14), 6553-6560 (2013).
- Huang, X. *et al.* Stretchable, wireless sensors and functional substrates for epidermal characterization of sweat. *Small.* **10** (15), 3083-3090 (2014).
- Rose, D. P. *et al.* Adhesive RFID Sensor Patch for Monitoring of Sweat Electrolytes. *IEEE Trans. Biomed. Eng.* **62** (6), 1457-1465 (2015).
- Gao, W. *et al.* Fully integrated wearable sensor arrays for multiplexed in situ perspiration analysis. *Nature.* **529** (7587), 509-514 (2016).
- Nilsson, G. E. Measurement of water exchange through skin. *Med. Biol. Eng. Comput.* **15** (3), 209-218 (1977).
- Tagami, H., Kobayashi, H., & Kikuchi, K. A portable device using a closed chamber system for measuring transepidermal water loss: comparison with the conventional method. *Skin Res. Technol.* **8** (1), 7-12 (2002).
- Nuutinen, J. *et al.* A closed unventilated chamber for the measurement of transepidermal water loss. *Skin Res. Technol.* **9** (2), 85-89 (2003).
- Imhof, R. E., De Jesus, M. E., Xiao, P., Ciorte, L. I., & Berg, E. P. Closed-chamber transepidermal water loss measurement: microclimate, calibration and performance. *Int. J. Cosmet. Sci.* **31** (2), 97-118 (2009).
- Sakaguchi, M., Ono, N., & Ohhashi, T. A new skin moisture meter using absolute hygrosensor. *Tech. Rep. IEICE.* **98** (309), 43-47 (1998).
- Sakaguchi, M. *et al.* Development of the new ventilation capsule type sweating-evaporation ratemeter - measurements of local sweating rates and evaporation rates. *Tech. Rep. IEICE.* **106** (253), 65-68 (2006).
- Ogai, K., Fukuoka, M., Kitamura, K., Uchida, K., & Nemoto, T. Development of a small wireless device for perspiration monitoring. *Med. Eng. Phys.* **38** (4), 391-397 (2016).
- Silicon Labs. *CP210x USB to UART Bridge VCP Drivers.* <http://www.silabs.com/products/mcu/Pages/USBtoUARTBridgeVCPDrivers.aspx> (2016).
- Claus, D., & Schondorf, R. Sympathetic skin response. In: *Recommendations for the Practice of Clinical Neurophysiology: Guidelines of the International Federation of Clinical Physiology (EEG Suppl. 52).* Deuschl, G., & Eisen, A., eds., Chapter 7.1, 277-282 (1999).
- Emad, M., Roshanzamir, S., Dabbaghmanesh, A., Ghasempoor, M. Z., & Eivazlou, H. Inclusion of Height and Limb Length when Interpreting Sympathetic Skin Response. *Iran J Med Sci.* **41** (1), 48-52 (2016).