

Video Article

Erratum: Larval RNA Interference in the Red Flour Beetle, *Tribolium castaneum*

URL: <https://www.jove.com/video/5754>

DOI: [doi:10.3791/5754](https://doi.org/10.3791/5754)

Keywords: Errata, Issue 104,

Date Published: 10/1/2015

Citation: Erratum: Larval RNA Interference in the Red Flour Beetle, *Tribolium castaneum*. *J. Vis. Exp.* (104), e5754, doi:10.3791/5754 (2015).

Abstract

An erratum was issued for [Larval RNA Interference in the Red Flour Beetle, *Tribolium castaneum*](#). There was a typo in the settings for Sutter P-87 in step 4.1.

Protocol

An erratum was issued for [Larval RNA Interference in the Red Flour Beetle, *Tribolium castaneum*](#). There was a typo in the settings for Sutter P-87 in step 4.1.

Step 4.1 was updated from:

Pull borosilicate glass needles (B100-50-15, O.D.: 1 mm, I.D.: 0.5 mm, 15 cm length) by a needle puller.

NOTE: For Sutter P-87 or P-97, use the setting "Heat = 70, Pull = 45, Vel = 75, Time = 90" or follow Chapter 2 of the Pipette Cookbook: Adherent Cell, *C. elegans*, *Drosophila*, & *Zebrafish*— Recommended Programs. The optimal settings vary depending on the model of the needle puller. The setting for a generic *Drosophila* injection needle is a good starting point.

to

Pull borosilicate glass needles (B100-50-15, O.D.: 1 mm, I.D.: 0.5 mm, 15 cm length) by a needle puller.

NOTE: For Sutter P-87 or P-97, use the setting "Heat = 770, Pull = 45, Vel = 75, Time = 90" or follow Chapter 2 of the Pipette Cookbook: Adherent Cell, *C. elegans*, *Drosophila*, & *Zebrafish*— Recommended Programs. The optimal settings vary depending on the model of the needle puller. The setting for a generic *Drosophila* injection needle is a good starting point.

Disclosures

No conflicts of interest declared.